

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-08062025-002

Gage ID PEA-22603 Gage S/N 22603 Description 10 lb. Test Weight Operating Procedure: OP-118 Unit of Meas. lbs Manufacturer Parker Cal. Date 8/6/2025 Next Due 8/6/2028 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 7129341906
Test weight as found 10.09 lbs.
Test weight as left 10.09 lbs.

APPROVED**By Kayla Myers at 2:36 pm, Aug 06, 2025**

Calibration due date for scale is on 4/11/2026
Scale is traceable to NIST by the following numbers: LA22-255, LA22-256, LA22-257

Calibrated By Taichi Daimo Signature  Date: 8/6/2025