

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	Manufacturer	Cut N Edge
Calibration Date	August 01, 2026	TIP Procedure	TIP-COILW01.1
Calibration Due Date	February 01, 2027	Calibration Technician	Freddy Seabourn
Serial Number	CES-105455	Order Number	81126
Instrument	10" AC Rectified Coil	Temperature / Humidity	68.5°F / RH 40.5%

CALIBRATION RESULTS

Polarity Check	Gauss Reading	Result
Positive	+102 G	PASS
Zero	0 G	—
Negative	-104 G	PASS

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	PASSED

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Gauss Meter, Alpha Lab Inc. GM-1 Digital Gauss Meter, S/N 3843 (Cert. Due 01/28/2027)
- TIP Master Hygrometer, S/N 31874 / 1H2048MCR / 130070751 (Cert. Due 01/28/2027)

Calibration Technician



Freddy Seabourn

Issue Date: August 01, 2026

APPROVED

By Kayla Myers at 9:57 am, Sep 14, 2026