

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04132026-014

<p>Gage ID PEA-0108253372 Gage S/N 0108253372 Description 12" Digital Caliper Operating Procedure: OP-110 Unit of Meas. Inch Manufacturer Insize Cal. Date 4/13/2026 Next Due 10/13/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID 1.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.999	Nominal	Maximum 1.001
Ref Type 81PC Gage Blocks	As Found 1.000	Accuracy	Fail Before No
	As Left 1.000	Accuracy	Fail After No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No.	81004
Std Due Date 1/26/2027			

Standard ID 3.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 2.999	Nominal	Maximum 3.001
Ref Type 81PC Gage Blocks	As Found 3.000	Accuracy	Fail Before No
	As Left 3.000	Accuracy	Fail After No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No.	81004
Std Due Date 1/26/2027			

Standard ID 5.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 4.999	Nominal	Maximum 5.001

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111
 Std Due Date 1/26/2027

Gage S/N 070111 NIST No. 81004

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No


Gage ID of Standard REF-070111
 Std Due Date 1/26/2027

Gage S/N 070111 NIST No. 81004

Standard ID	9.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	81PC Gage Blocks	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111
 Std Due Date 1/26/2027

Gage S/N 070111 NIST No. 81004

Calibrated By Chad Adams Signature  Date: 4/13/2026

APPROVED
 By Kayla Myers at 2:20 pm, May 28, 2026