

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-01162026-014

<p>Gage ID PEA-1608211172 Gage S/N 1608211172 Description 12" Digital Caliper Operating Procedure: OP-110 Unit of Meas. inch Manufacturer insize Cal. Date 1/16/2026 Next Due 7/16/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15392 Job 15125

Standard ID 1.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.999	Nominal 1.000	Maximum 1.001
Ref Type 81PC Gage Blocks	As Found 1.000	Accuracy 0.000	Fail Before No
	As Left 1.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No. 81004	
Std Due Date 1/26/2027			

Standard ID 3.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 2.999	Nominal 3.000	Maximum 3.001
Ref Type 81PC Gage Blocks	As Found 3.000	Accuracy 0.000	Fail Before No
	As Left 3.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No. 81004	
Std Due Date 1/26/2027			

Standard ID 5.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 4.999	Nominal 5.000	Maximum 5.001

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No


Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Standard ID	9.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	81PC Gage Blocks	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Calibrated By Chad Adams Signature  Date: 1/16/2026

APPROVED

By Kayla Myers at 11:14 am, Jan 20, 2026