

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-01062025-000

Gage ID PEA-2306211150	Environmental Conditions
Gage S/N 2306211150	Temperature 68 Deg F +/- 2
Description 12" Digital Depth Gauge	Humidity 35-55%
Operating Procedure: OP-123	
Unit of Meas. inch	Approved Yes
Manufacturer insize	Customer Info. Peak NDT Solutions
Cal. Date 1/6/2025	
Next Due 7/6/2025	
Cal. Freq. 6.00 Months	
Location Lab	

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13719 Job 11839

Standard ID 1.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.999	Nominal 1.000	Maximum 1.001
Ref Type 81PC Gage Blocks	As Found 1.000	Accuracy 0.000	Fail Before No
	As Left 1.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-070111			
Std Due Date 1/26/2027	Gage S/N 070111	NIST No. 81004	
Standard ID 3.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 2.999	Nominal 3.000	Maximum 3.001
Ref Type 81PC Gage Blocks	As Found 3.000	Accuracy 0.000	Fail Before No
	As Left 3.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-070111			
Std Due Date 1/26/2027	Gage S/N 070111	NIST No. 81004	
Standard ID 5.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 4.999	Nominal 5.000	Maximum 5.001

Ref Type	81PC Gage Blocks	As Found	5.001	Accuracy	0.001	Fail Before	No
		As Left	5.001	Accuracy	0.001	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	9.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	81PC Gage Blocks	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Calibrated By Chad AdamsSignature Date: 1/6/2025**APPROVED***By Kayla Myers at 4:26 pm, Jan 07, 2025*