

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-01162026-010

<p>Gage ID PEA-8235-1 Gage S/N 8235-1 Description 50-0-50 gauss meter Operating Procedure: OP-117 Unit of Meas. gauss Manufacturer R.B. Annis Cal. Date 1/16/2026 Next Due 7/16/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes</p> <p>Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15392 Job 15125

Standard ID	+25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26
Ref Type	50-0-50 Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-01						
Std Due Date	8/21/2026	Gage S/N	TSPSTD-01	NIST No.	171843		

Standard ID	+50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	50-0-50 Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-01						
Std Due Date	8/21/2026	Gage S/N	TSPSTD-01	NIST No.	171843		

Standard ID	-25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26

Ref Type	50-0-50 Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No


Gage ID of Standard REF-TSPSTD-01

Std Due Date	8/21/2026	Gage S/N	TSPSTD-01	NIST No.	171843
--------------	-----------	----------	-----------	----------	--------

Standard ID	-50	Uncertainty	Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum 51
Ref Type	50-0-50 Gauss Meter	As Found	50	Accuracy	0	Fail Before No
		As Left	50	Accuracy	0	Fail After No

Gage ID of Standard REF-TSPSTD-01

Std Due Date	8/21/2026	Gage S/N	TSPSTD-01	NIST No.	171843
--------------	-----------	----------	-----------	----------	--------

Calibrated By Chad Adams Signature  Date: 1/16/2026

APPROVED
By Kayla Myers at 11:12 am, Jan 20, 2026