

## TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706  
Phone: (432) 296-4992 Email: Chad@tspndt.com

### Certificate of Calibration

Certificate No TSPW-02132025-001

<b>Gage ID</b> PEA-8235-1 <b>Gage S/N</b> 8235-1 <b>Description</b> 50-0-50 gauss meter <b>Operating Procedure:</b> OP-117 <b>Unit of Meas.</b> gauss <b>Manufacturer</b> R.B. Annis <b>Cal. Date</b> 2/13/2025 <b>Next Due</b> 8/13/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> Temperature 68 Deg F +/- 2 Humidity 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

#### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### Findings

SO 13910 Job 13190

<b>Standard ID</b>	+25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26
<b>Ref Type</b>	50-0-50 Gauss Meter	<b>As Found</b>	25	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	25	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-TSPSTD-01						
<b>Std Due Date</b>	8/8/2025	<b>Gage S/N</b>	TSPSTD-01	<b>NIST No.</b>	154783		
<b>Standard ID</b>	+50	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	49	<b>Nominal</b>	50	<b>Maximum</b>	51
<b>Ref Type</b>	50-0-50 Gauss Meter	<b>As Found</b>	50	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	50	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-TSPSTD-01						
<b>Std Due Date</b>	8/8/2025	<b>Gage S/N</b>	TSPSTD-01	<b>NIST No.</b>	154783		
<b>Standard ID</b>	-25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26

Ref Type	50-0-50 Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-01

Std Due Date 8/8/2025

Gage S/N TSPSTD-01 NIST No. 154783

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	50-0-50 Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-01

Std Due Date 8/8/2025

Gage S/N TSPSTD-01 NIST No. 154783

Calibrated By Chad Adams

Signature



Date: 2/13/2025

**APPROVED***By Kayla Myers at 11:19 am, Mar 03, 2025*