

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No (None)

Gage ID PEA-SKF112 Gage S/N SKF112 Description Lead Gauge Operating Procedure: OP-124 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 9/23/2025 Next Due 3/23/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Contact point spacing at 1"
Target: 0.0040 As Found: 0.0040 As Left: 0.0040

Contact point spacing at 2"
Target: 0.0080 As Found: 0.0080 As Left: 0.0080

Tolerance: +/- .0020"

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: LS1007 Lead Gauge Standard - Serial Number: JU06FM0027

NIST Number: Doc#301641 - Calibration Due Date: 6/17/26

Reference Standard: LS1008 Lead Gauge Standard - Serial Number: RG15DM0010

NIST Number: Doc#297680 - Calibration Due Date: 6/17/26

Calibrated By Chad Adams

Signature

Date: 9/23/2025**APPROVED****By Kayla Myers at 7:28 am, Sep 25, 2025**