

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-01162026-009

<p>Gage ID PEA-CES-105455 Gage S/N CES-105455 Description 10-1/2" Magnetic Coil Operating Procedure: OP-116 Unit of Meas. gauss Manufacturer Cut-N-Edge Cal. Date 1/16/2026 Next Due 7/16/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15392 Job 15125

As Found: Passed As Left: Passed

Coil putting out magnetic field from both the positive and negative side equal to or greater than 50 gauss for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: 50-0-50 Magnetometer - Serial Number: TSPSTD-01

NIST Number: Doc#171843 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams Signature  Date: 1/16/2026

APPROVED
By Kayla Myers at 11:11 am, Jan 20, 2026