

**TSP Calibration, Inc.**

1110 S. FM 1788 Suite 38 Midland, TX 79706  
Phone: (432) 296-4992 Email: Chad@tspndt.com

**Certificate of Calibration**

Certificate No TSPW-07082025-021

<b>Gage ID</b> PEA-CES-105455 <b>Gage S/N</b> CES-105455 <b>Description</b> 10 1/2" Magnetic Coil <b>Operating Procedure:</b> OP-116 <b>Unit of Meas.</b> gauss <b>Manufacturer</b> Cut-N-Edge <b>Cal. Date</b> 7/8/2025 <b>Next Due</b> 1/8/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 14534 Job 14212

AS FOUND: PASSED AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. TSPSTD-01 was used for this calibration.

Calibration due date for gauss meter is on 8/8/2025

Gauss meter is traceable to NIST by the following number: Certification #154783

Calibrated By Chad Adams

Signature

Date: 7/8/2025**APPROVED****By Kayla Myers at 8:32 am, Aug 15, 2025**