# TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

## Certificate of Calibration

#### Certificate No TSPW-02132025-003

Gage ID PEA-25437

Gage S/N 25437

Description 10 lb Test Weight

Operating

OP-118

Procedure:

Unit of Meas. pounds

Manufacturer Parker

Cal. Date 2/13/2025 Next Due 2/13/2026

Cal. Freq. 1.00

Location Lab

**Environmental Conditions** 

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### **Findings**

SO 13910 Job 13190

AS FOUND: PASSED

AS LEFT: PASSED

Years

Using calibrated scale serial no. 8335029565

Test weight as found 10.110 lbs.

Test weight as left 10.110 lbs.

Calibration due date for scale is on 8/8/2025

Scale is traceable to NIST by the following numbers: Doc# 154780

Calibrated By Chad Adams

Signature ,

Date:

2/13/2025

**APPROVED** 

By Kayla Myers at 11:20 am, Mar 03, 2025