

# TSP Calibration, Inc.

1110 S. FM 1788 #42  
 Phone: (432) 296-4992 Email: Chad@tspndt.com  
 Midland, TX 79706

## Certificate of Calibration

Certificate No TSPW-01162026-013

<p><b>Gage ID</b> PEA-2419001777  <b>Gage S/N</b> 2419001777  <b>Description</b> UV Light Meter  <b>Operating Procedure:</b> OP-119  <b>Unit of Meas.</b> Microwatts/cm2  <b>Manufacturer</b> General  <b>Cal. Date</b> 1/16/2026  <b>Next Due</b> 7/16/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 Deg F +/- 2                  Humidity 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

SO 15392 Job 15125

We certify the equipment used for this calibration is traceable to NIST through the following number(s):  
 Reference Standard: UV Light - Serial Number: 1911981  
 NIST Number: Doc#171852 - Calibration Due Date: 8/21/2026

<b>Standard ID</b>	12"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	2891.8	<b>Nominal</b>	3044	<b>Maximum</b>	3196.2
<b>Ref Type</b>	Light Meter	<b>As Found</b>	3108	<b>Accuracy</b>	64	<b>Fail Before</b>	No
		<b>As Left</b>	3108	<b>Accuracy</b>	64	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-1705427						
<b>Std Due Date</b>	8/21/2026	<b>Gage S/N</b>	1705427	<b>NIST No.</b>	171854		

<b>Standard ID</b>	15"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	2088.1	<b>Nominal</b>	2198	<b>Maximum</b>	2307.9
<b>Ref Type</b>	Light Meter	<b>As Found</b>	2244	<b>Accuracy</b>	46	<b>Fail Before</b>	No
		<b>As Left</b>	2244	<b>Accuracy</b>	46	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-1705427						
<b>Std Due Date</b>	8/21/2026	<b>Gage S/N</b>	1705427	<b>NIST No.</b>	171854		

Calibrated By Chad Adams

Signature



Date: 1/16/2026

**APPROVED**

*By Kayla Myers at 11:13 am, Jan 20, 2026*