

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-07082025-017

Gage ID PEA-2419001777 Gage S/N 2419001777 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 7/8/2025 Next Due 1/8/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14534 Job 14212

Calibrated black light serial no. 1911981 was used during this calibration.

Calibration due date for black light is on 8/8/2025

Black light is traceable to NIST by the following numbers: Doc# 154779

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2935.5	Nominal	3090	Maximum	3244.5
Ref Type	Light Meter	As Found	3113	Accuracy	23	Fail Before	No
		As Left	3113	Accuracy	23	Fail After	No
Gage ID of Standard	REF-1705427						
Std Due Date	8/12/2025	Gage S/N	1705427	NIST No.	154789		

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2198.3	Nominal	2314	Maximum	2429.7
Ref Type	Light Meter	As Found	2353	Accuracy	39	Fail Before	No
		As Left	2353	Accuracy	39	Fail After	No
Gage ID of Standard	REF-1705427						
Std Due Date	8/12/2025	Gage S/N	1705427	NIST No.	154789		

Calibrated By Chad Adams

Signature 

Date: 7/8/2025

APPROVED
By Kayla Myers at 8:41 am, Aug 15, 2025