

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-02232026-005

<p>Gage ID PEA-4055 13 Gage S/N 4055 13 Description 5-step block Operating Procedure: OP-126 Unit of Meas. inch Manufacturer Curtis Industries Cal. Date 2/23/2026 Next Due 2/23/2027 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15584 Job 15174

Standard ID .100 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.098	Nominal	0.100	Maximum 0.102
Ref Type 12" Digital Caliper	As Found	0.100	Accuracy	0.000	Fail Before No
	As Left	0.100	Accuracy	0.000	Fail After No
Gage ID of Standard REF-1095377					
Std Due Date 8/25/2026	Gage S/N 1095377		NIST No. 171845		

Standard ID .200 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.198	Nominal	0.200	Maximum 0.202
Ref Type 12" Digital Caliper	As Found	0.200	Accuracy	0.000	Fail Before No
	As Left	0.200	Accuracy	0.000	Fail After No
Gage ID of Standard REF-1095377					
Std Due Date 8/25/2026	Gage S/N 1095377		NIST No. 171845		

Standard ID .300 "	Uncertainty		Units		Type V
Limited Use? No	Minimum	0.298	Nominal	0.300	Maximum 0.302

Ref Type	5 Step Block	As Found	0.291	Accuracy	-0.009	Fail Before	Yes
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.387	Accuracy	-0.013	Fail Before	Yes
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.494	Accuracy	-0.006	Fail Before	Yes
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Calibrated By Chad Adams Signature  Date: 2/23/2026

APPROVED
By Kayla Myers at 9:08 pm, Jun 02, 2026