

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08252025-012

Gage ID PEA-21436 Gage S/N 21436 Description A/C Yoke Operating Procedure: OP-115 Unit of Meas. pounds Manufacturer Electro-Spect Cal. Date 8/25/2025 Next Due 2/25/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14759 Job 14274

AS FOUND: PASSED AS LEFT: PASSED

A/C Yoke picked up and held required 10 lb. test weight for calibration.
Calibrated test weight serial no. 3452 was used for this calibration.
Test weight calibration is due on 8/21/2026
Test weight is traceable to NIST by the following number: Doc#171848

Calibrated By Chad Adams

Signature

Date: 8/25/2025**APPROVED****By Kayla Myers at 10:51 am, Aug 26, 2025**