



Tubular Inspection Products Inc.

Certificate of Calibration

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017/ISO 9001/API Q1/7-2

Customer: **Peak NDT Solutions, LLC**

Calibration Date: **July 09, 2026**

Calibration Due Date: **January 09, 2027**

Serial Number: **3001265176**

Instrument: **12" Digital Caliper**

Manufacturer: **Insize**

TIP Procedure: **TIP-CALW01.1**

Calibration Technician: **Ryan Britton**

Order Number: **80897**

Results

	Reference Size	Actual Measurement	Deviation	Tolerance	Result
<u>Step</u>	<u>1.000"</u>	<u>1.0000"</u>	<u>0.0000"</u>	<u>+/-0.001"</u>	<u>Pass</u>
<u>Step</u>	<u>2.000"</u>	<u>2.0000"</u>	<u>0.0000"</u>	<u>+/-0.001"</u>	<u>Pass</u>
<u>Step</u>	<u>3.000"</u>	<u>3.0000"</u>	<u>0.0000"</u>	<u>+/-0.001"</u>	<u>Pass</u>
<u>Step</u>	<u>4.000"</u>	<u>4.0000"</u>	<u>0.0000"</u>	<u>+/-0.001"</u>	<u>Pass</u>

<u>Instrument Received</u>	<u>Instrument Returned</u>	<u>Instrument Status</u>
In Tolerance	In Tolerance	Passed

APPROVED

By Kayla Myers at 10:39 am, Jul 24, 2026

Temperature/Humidity: 68.5°F / RH 40.5%

Issue Date for Certificate: **July 7, 2026**

TIP Master Optical Comparator, Keyence IM-8030,
4C0Q000412 Cal Date: 06-26-2027

TIP Masters Blocks : 13232,14002,14340,13207 Cal Date: 01-25-2027

TIP Master Hygrometer SN: 31874/1H2048MCR, 130070751, Cal Date: 01-29-27

Approved By: 