Gage ID: PEA-7632

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08252025-006

Gage ID PEA-7632

Gage S/N 7632

Description 20-0-20 gauss meter

Operating

OP-117

Procedure:

Unit of Meas. gauss

Manufacturer R.B. Annis

Cal. Date 8/25/2025 **Next Due** 2/25/2026

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14759 Job 14274

Standard ID	+10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Ref Type		0 Gauss Meter	As Found		10	Accuracy		0	Fail Before	No	
Rei Type	2002		As Left		10	Accuracy		0	Fail After	No	
Gage ID of Stan	dard	REF-TSPSTD-02									
Std Due Date		8/21/2026	Gage S/N	TSPSTD-02	I	NIST No.	171844				
	.00		Uncertainty			Units			Туре	V	
Standard ID	+20				10	Nominal		20	Maximum		
Limited Use?	No		Minimum		19						
Ref Type	20-0-2	0 Gauss Meter	As Found		20	Accuracy		0	Fail Before		
			As Left		20	Accuracy		0	Fail After	No	
Gage ID of Stan	dard	REF-TSPSTD-02									
Std Due Date		8/21/2026	Gage S/N	TSPSTD-02	SPSTD-02 NIST No.		171844				
Ot do d ID	-10		Uncertainty			Units			Туре	V	
Standard ID Limited Use?	No		Minimum		9	Nominal		10	Maximum		

Gage ID: PEA-7632

Page

2 of 2

Ref Type

20-0-20 Gauss Meter

As Found

Accuracy 10

Fail Before No

As Left

10 Accuracy Fail After No

Gage ID of Standard

REF-TSPSTD-02

Std Due Date

8/21/2026

Gage S/N TSPSTD-02

NIST No.

171844

Standard ID

-20

Uncertainty

Units 19 Nominal

Type V Maximum 21 20

Limited Use? Ref Type

No 20-0-20 Gauss Meter

Minimum As Found

As Left

Accuracy 20 20 Accuracy 0 Fail Before No Fail After No

Gage ID of Standard

REF-TSPSTD-02

Std Due Date

8/21/2026

Gage S/N TSPSTD-02

NIST No.

171844

Calibrated By Chad Adams

Signature

Date:

8/25/2025

APPROVED

By Kayla Myers at 10:47 am, Aug 26, 2025