

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08122024-019

<p>Gage ID PEA-1011-2 Gage S/N 1011-2 Description Profile Gauge Operating Procedure: Visual Unit of Meas. Manufacturer PMC Lone Star Cal. Date 8/12/2024 Next Due 8/12/2027 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13176 Job 13190

Flatness- PASS
Damage- PASS
Wear- PASS

As Found: Passed
As Left: Passed

APPROVED

By Kayla Myers at 8:39 am, Aug 13, 2024

Calibrated By Chad Adams

Signature 

Date: 8/12/2024

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08122024-020

<p>Gage ID PEA-1011-3 Gage S/N 1011-3 Description Profile Gauge Operating Procedure: Visual Unit of Meas. Manufacturer PMC Lone Star Cal. Date 8/12/2024 Next Due 8/12/2027 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: right;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13176 Job 13190

Flatness- PASS
Damage- PASS
Wear- PASS

As Found: Passed
As Left: Passed

APPROVED

By Kayla Myers at 8:39 am, Aug 13, 2024

Calibrated By Chad Adams

Signature



Date: 8/12/2024

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08122024-010

<p>Gage ID PEA-1011-4 Gage S/N 1011-4 Description Profile Gauge Operating Procedure: Visual Unit of Meas. Manufacturer PMC Lone Star Cal. Date 8/12/2024 Next Due 8/12/2027 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: right;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13176 Job 13190

Flatness- PASS
Damage- PASS
Wear- PASS

As Found: Passed
As Left: Passed

APPROVED
By Kayla Myers at 8:40 am, Aug 13, 2024

Calibrated By Chad Adams Signature  Date: 8/12/2024

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08122024-011

<p>Gage ID PEA-1011-1 Gage S/N 1011-1 Description Profile Gauge Operating Procedure: Visual Unit of Meas. Manufacturer PMC Lone Star Cal. Date 8/12/2024 Next Due 8/12/2027 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: right;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 13176 Job 13190

Flatness- PASS
Damage- PASS
Wear- PASS

As Found: Passed
As Left: Passed

APPROVED
By Kayla Myers at 8:40 am, Aug 13, 2024

Calibrated By Chad Adams Signature  Date: 8/12/2024