# Tech Service Products

**1110 S. FM 1788 #42**
**Phone: (432) 296-4992 Email: Chad@tspndt.com**
**Midland,TX 79706**

## Certificate of Calibration

**Certificate No TSPW-10272026-007**

| | |
|---|---|
| **Gage ID** | PEA-33039 |
| **Gage S/N** | 33039 |
| **Description** | 10 lb Test Weight |
| **Operating Procedure:** | OP-118 |
| **Unit of Meas.** | Pounds |
| **Manufacturer** | Parker |
| **Cal. Date** | 10/27/2026 |
| **Next Due** | 10/27/2029 |
| **Cal. Freq.** | 3.00 Years |
| **Location** | Lab |

**Environmental Conditions**

| | |
|---|---|
| **Temperature** | 68 Deg F +/- 2 |
| **Humidity** | 20-55% |

| | |
|---|---|
| **Approved** | Yes |
| **Customer Info.** | Peak NDT Solutions |

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Order No. 1067364 Job No. 261027-2

As Found: Passed As Left: Passed

| Weight as Found | Weight as Left |
|---|---|
| 10.095 lbs. | 10.095 lbs. |

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: Digital Scale - Serial Number: 8335029565
NIST Numbers: 187824 - Calibration Due Date: 8/27/2027

**Calibrated By** Chad Adams **Signature**  **Date:** 10/27/2026

***APPROVED***
***By Kayla Myers at 11:51 am, Oct 29, 2026***