Gage ID: PEA-3960

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08252025-005

Gage ID PEA-3960

Gage S/N 3960

Description 10 lb Test Weight

Operating OP-118

Procedure:

Unit of Meas. pounds

Manufacturer Electro-Spect

Cal. Date 8/25/2025 Next Due 8/25/2028

Cal. Freq. 3.00

Location Lab

Years

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology

Findings

SO 14759 Job 14274

AS FOUND: PASSED

AS LEFT: PASSED

Using calibrated scale serial no. 8335029565

Test weight as found 10.040 lbs.

Test weight as left 10.040 lbs.

Calibration due date for scale is on 8/21/2026

Scale is traceable to NIST by the following numbers: Doc#171842

Calibrated By Chad Adams

Signature

Date:

8/25/2025

APPROVED

By Kayla Myers at 10:47 am, Aug 26, 2025