

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-02232026-004

<p>Gage ID PEA-20064 Gage S/N 20064 Description UT Thickness Gauge Operating Procedure: OP-121 Unit of Meas. inch Manufacturer Dakota Cal. Date 2/23/2026 Next Due 8/23/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15584 Job 15174

Standard ID	.100 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	5 Step Block	As Found	0.088	Accuracy	-0.012	Fail Before	Yes
		As Left	0.100	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-337214						
Std Due Date	8/21/2026	Gage S/N	337214	NIST No.	171846		

Standard ID	.200 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	5 Step Block	As Found	0.189	Accuracy	-0.011	Fail Before	Yes
		As Left	0.200	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-337214						
Std Due Date	8/21/2026	Gage S/N	337214	NIST No.	171846		

Standard ID	.300 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Ref Type	12" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/25/2026

Gage S/N 1095377

NIST No. 171845

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	12" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/25/2026

Gage S/N 1095377

NIST No. 171845

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	12" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377


Std Due Date 8/25/2026

Gage S/N 1095377

NIST No. 171845

Calibrated By Chad Adams

Signature



Date: 2/23/2026

APPROVED

By Kayla Myers at 9:09 pm, Jun 02, 2026