

## TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706  
 Phone: (432) 296-4992 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-08252025-009

<b>Gage ID</b> PEA-20064 <b>Gage S/N</b> 20064 <b>Description</b> UT Thickness Gauge <b>Operating Procedure:</b> OP-121 <b>Unit of Meas.</b> inch <b>Manufacturer</b> Dakota <b>Cal. Date</b> 8/25/2025 <b>Next Due</b> 2/25/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

## Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

## Findings

SO 14759 Job 14274

Standard ID	Uncertainty	Units	Type	V
.100 "	Minimum	Nominal	Maximum	0.102
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 5 Step Block	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No. 171846		
Std Due Date 8/21/2026				
Standard ID	Uncertainty	Units	Type	V
.200 "	Minimum	Nominal	Maximum	0.202
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 5 Step Block	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No. 171846		
Std Due Date 8/21/2026				
Standard ID	Uncertainty	Units	Type	V
.300 "	Minimum	Nominal	Maximum	0.302
Limited Use? No				

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846


Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846

Calibrated By	Chad Adams	Signature		Date:	8/25/2025
---------------	------------	-----------	--	-------	-----------

**APPROVED**

By Kayla Myers at 10:50 am, Aug 26, 2025