

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-06092026-004

Gage ID PEA-0704213202 Gage S/N 0704213202 Description 8" Digital Depth Gauge Operating Procedure: OP-123 Unit of Meas. Inch Manufacturer Insize Cal. Date 6/9/2026 Next Due 12/9/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No.: 1056500 Job: 15456

Standard ID	1.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.999	Nominal	1.000	Maximum	1.001
Ref Type	81PC Gage Blocks	As Found	0.999	Accuracy	-0.001	Fail Before	No
		As Left	0.999	Accuracy	-0.001	Fail After	No
Gage ID of Standard	REF-070111						
Std Due Date	1/26/2027	Gage S/N	070111	NIST No.	81004		

Standard ID	3.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2.999	Nominal	3.000	Maximum	3.001
Ref Type	81PC Gage Blocks	As Found	2.999	Accuracy	-0.001	Fail Before	No
		As Left	2.999	Accuracy	-0.001	Fail After	No
Gage ID of Standard	REF-070111						
Std Due Date	1/26/2027	Gage S/N	070111	NIST No.	81004		

Standard ID	5.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4.999	Nominal	5.000	Maximum	5.001

Ref Type	81PC Gage Blocks	As Found	4.999	Accuracy	-0.001	Fail Before	No
		As Left	4.999	Accuracy	-0.001	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	6.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	5.999	Nominal	6.000	Maximum	6.001
Ref Type	81PC Gage Blocks	As Found	5.999	Accuracy	-0.001	Fail Before	No
		As Left	5.999	Accuracy	-0.001	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	6.999	Accuracy	-0.001	Fail Before	No
		As Left	6.999	Accuracy	-0.001	Fail After	No

Gage ID of Standard REF-070111


Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Calibrated By Chad Adams

Signature



Date: 6/9/2026

APPROVED

By Kayla Myers at 4:18 pm, Jul 22, 2026