

TSP Calibration, Inc.

1110 S. FM 1788 #42

Phone: (432) 296-4992 Email: Chad@tspndt.com

Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-06092026-006

Gage ID PEA-2037714 Gage S/N 2037714 Sensor No. 2037717 Description Dual Light Meter Operating Procedure: OP-119/OP-130 Unit of Meas. Microwatts/cm2 & Foot Candles Manufacturer Spectroline Cal. Date 6/9/2026 Next Due 12/9/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No.: 1056500 Job: 15456

Visible Light Readings:	Master	As Found
High:	104.8	107.1
Low:	19.3	20.6

NOTE: If only an "AS FOUND" reading is stated, "AS FOUND" and "AS LEFT" are the same with no adjustments.

Tolerance: +/-5% or +/-2 foot candles, whichever is greater.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: UV Light - Serial Number: 1911981

NIST Number: Doc#171852 - Calibration Due Date: 8/21/2026

Reference Standard: Dual Light Meter - Serial Number: 1705427 Sensor: 1705430

NIST Number: Doc#171854 - Calibration Due Date: 8/21/2026

UV Light Readings

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2957.35	Nominal	3113	Maximum	3268.65
Ref Type	Light Meter	As Found	3141	Accuracy	28	Fail Before	No
		As Left	3141	Accuracy	28	Fail After	No

Gage ID of Standard REF-1705427

Std Due Date 8/21/2026

Gage S/N 1705427

NIST No. 171854

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2155.55	Nominal	2269	Maximum	2382.45
Ref Type	Light Meter	As Found	2288	Accuracy	19	Fail Before	No
		As Left	2288	Accuracy	19	Fail After	No

Gage ID of Standard REF-1705427


Std Due Date 8/21/2026

Gage S/N 1705427

NIST No. 171854

Calibrated By Chad Adams

Signature



Date: 6/9/2026

APPROVED*By Kayla Myers at 4:19 pm, Jul 22, 2026*