

Company:	Peak NDT Solutions	Doc #:	189896	Print#:	
Address:	515A Verot School Road	OEM #:		Serial#:	SDM-3636
	Lafayette LA	Contact:		Location:	Calibration Lab
Zip:	70506	Dept:	WO# 100926-09	Mfg.:	Magnaflux
Page:	1 of 1	Cal Date:	10/12/2026	Model:	Yoke
		PO#:		Gage:	115VAC, 60Hz, 4 Amps

Yoke Will Pick Up A 10 Pound Weight With Poles Spaced 2-4"

Master	Actual	Deviation	Serial	Tol
+10.000000	+10.000000	+0.000000		+0.000000 / 0.000000

Comments

Procedure# SDM-20 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 48XL

Weight Set 2oz-50 lbs, Rice Lake, Class "F" NIST

Traceable to NIST via NAMAS certification no: Cert# 25-334, 25-376,
25-335

Last / Next Cal Dates 05/05/2025 → 05/05/2028

Gage Status PASS**Next Calibration Due: 04/12/2027**

Certified By: Mike Thibeaux

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND READING IS STATED" "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibration certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

Signature

**APPROVED****By Kayla Myers at 3:28 pm, Oct 13, 2026**