

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-06262025-015

Gage ID PEA-2905231617
 Gage S/N 2905231617
 Description 8" Digital Caliper
 Operating Procedure: OP-110
 Unit of Meas. inch
 Manufacturer insize
 Cal. Date 6/26/2025
 Next Due 12/26/2025
 Cal. Freq. 6.00 Months
 Location Lab

Environmental Conditions

Temperature 68 Deg F +/- 2
 Humidity 35-55%

Approved Yes
 Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14497 Job 14202

Standard ID	Uncertainty	Units	Type	V
1.000 "	Minimum	Nominal	Maximum	1.001
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 81PC Gage Blocks	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No. 81004		
Std Due Date 1/26/2027				
Standard ID	Uncertainty	Units	Type	V
3.000 "	Minimum	Nominal	Maximum	3.001
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 81PC Gage Blocks	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-070111	Gage S/N 070111	NIST No. 81004		
Std Due Date 1/26/2027				
Standard ID	Uncertainty	Units	Type	V
5.000 "	Minimum	Nominal	Maximum	5.001
Limited Use? No				

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Standard ID	6.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	5.999	Nominal	6.000	Maximum	6.001
Ref Type	81PC Gage Blocks	As Found	6.000	Accuracy	0.000	Fail Before	No
		As Left	6.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Calibrated By	Chad Adams	Signature		Date:	6/26/2025
---------------	------------	-----------	--	-------	-----------

APPROVED

By Kayla Myers at 11:30 am, Jun 27, 2025