

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company:	Peak NDT Solutions	Doc #:	189885	Print#:	
Address:	515A Verot School Road Lafayette LA	OEM #:		Serial#:	0429
Zip:	70506	Contact:		Location:	Calibration Lab
Page:	1 of 1	Dept:	WO# 100926-09	Mfg.:	Drilco
		Cal Date:	10/12/2026	Model:	DC Coil
		PO#:		Gage:	DC Coil

Comments

Procedure# SDM-17 Rev.1

NOTE DC Coil went through a verification test using a +/- gauss magnetometer to insure that the unit functions properly.

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 15540250

Multimeter, Fluke, Model 87V

Traceable to NIST via NAMAS certification no: Cert# 144311

Last / Next Cal Dates 12/06/2023 → 12/06/2026

Gage Status PASS

Next Calibration Due: 04/12/2027


Certified By: Mike Thibeaux

This certificate is not valid unless all 1 page(s) are present.

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibration certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

Signature 

APPROVED

By Kayla Myers at 3:34 pm, Oct 13, 2026