

## TSP Calibration, Inc.

1110 S. FM 1788 #42  
 Phone: (432) 296-4992 Email: Chad@tspndt.com  
 Midland, TX 79706

## Certificate of Calibration

Certificate No TSPW-12112025-013

<b>Gage ID</b> PEA-0704213202 <b>Gage S/N</b> 0704213202 <b>Description</b> 8" Digital Depth Gauge <b>Operating Procedure:</b> OP-123 <b>Unit of Meas.</b> inch <b>Manufacturer</b> insize <b>Cal. Date</b> 12/11/2025 <b>Next Due</b> 6/11/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 25-45%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
---	---

## Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

## Findings

SO 15240 Job 14897

<b>Standard ID</b>	1.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.999	<b>Nominal</b>	1.000	<b>Maximum</b>	1.001
<b>Ref Type</b>	81PC Gage Blocks	<b>As Found</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-070111						
<b>Std Due Date</b>	1/26/2027	<b>Gage S/N</b>	070111	<b>NIST No.</b>	81004		
<b>Standard ID</b>	3.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	2.999	<b>Nominal</b>	3.000	<b>Maximum</b>	3.001
<b>Ref Type</b>	81PC Gage Blocks	<b>As Found</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-070111						
<b>Std Due Date</b>	1/26/2027	<b>Gage S/N</b>	070111	<b>NIST No.</b>	81004		
<b>Standard ID</b>	5.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	4.999	<b>Nominal</b>	5.000	<b>Maximum</b>	5.001

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Standard ID	6.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	5.999	Nominal	6.000	Maximum	6.001
Ref Type	81PC Gage Blocks	As Found	6.000	Accuracy	0.000	Fail Before	No
		As Left	6.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004
Std Due Date	1/26/2027				

Calibrated By	Chad Adams	Signature		Date:	12/11/2025
---------------	------------	-----------	--	-------	------------

**APPROVED***By Kayla Myers at 12:36 pm, Dec 17, 2025*