Gage ID: PEA-11212

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-06262025-012

Gage ID PEA-11212

Gage S/N 11212

Description 50-0-50 gauss meter

Operating OP-117

Procedure:

Unit of Meas. gauss

Manufacturer R.B. Annis

Cal. Date 6/26/2025 Next Due 12/26/2025

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Job 14202 SO 14497

Standard ID Limited Use? Ref Type	+25 No 50-0-5	0 Gauss Meter	Uncertainty Minimum As Found As Left		24 25 25	Units Nominal Accuracy Accuracy		25 0 0	Type Maximum Fail Before Fail After	26 No	
Gage ID of Stan Std Due Date	dard	REF-TSPSTD-01 8/8/2025	Gage S/N	TSPSTD-01		NIST No.	154783				
Standard ID	+50		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		49	Nominal		50	Maximum	51	
Ref Type		0 Gauss Meter	As Found		50	Accuracy		0	Fail Before	No	
Kerrype	30 0 0	, o Gauss Meter	As Left		50	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	REF-TSPSTD-01	Gage S/N	TSPSTD-01							
Std Due Date		8/8/2025			NIST No.		154783				
Standard ID	-25		Uncertainty			Units		1003 1500	Туре		
Limited Use?	No		Minimum		24	Nominal		25	Maximum	26	

Gage ID: PEA-11212

Page

2 of 2

Ref Type

50-0-50 Gauss Meter

As Found

25 Accuracy 0 Fail Before No

Gage ID of Standard

As Left

25 Accuracy

0 Fail After No

REF-TSPSTD-01

Std Due Date

8/8/2025

Gage S/N TSPSTD-01

NIST No.

154783

Standard ID -50

No

50-0-50 Gauss Meter

Uncertainty

49 Nominal

Units

Type ∨ 50 Maximum 51

Limited Use? Ref Type

Minimum As Found As Left

50 Accuracy 50 Accuracy 0 Fail Before No

Fail After No

Gage ID of Standard

REF-TSPSTD-01

Std Due Date

8/8/2025

Gage S/N TSPSTD-01

NIST No.

154783

Calibrated By

Chad Adams

Signature

Date:

6/26/2025

APPROVED

By Kayla Myers at 11:32 am, Jun 27, 2025