

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-12112025-006

Gage ID PEA-51504 Gage S/N 51504 Description IR Thermometer Operating Procedure: OP-200 Unit of Meas. Degrees Fahrenheit Manufacturer N/A Cal. Date 12/11/2025 Next Due 12/11/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 25-45% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15240 Job 14897

As Found: Passed As Left: Passed

Heat Setting 1 Master: 210 As Found: 218

Heat Setting 2 Master: 365 As Found: 356

Heat Setting 3 Master: 528 As Found: 521

Heat Setting 4 Master: 667 As Found: 654

Heat Setting 5 Master: 728 As Found: 733

Tolerance: +/- 7.5%

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Fluke 62 Max IR Thermometer - Serial Number: 44070254WS

NIST Number: Doc#171847 - Calibration Due Date: 6/17/26

Measurements were achieved using a heating element at 5 different settings.

APPROVED**By Kayla Myers at 12:39 pm, Dec 17, 2025**Calibrated By Chad AdamsSignature Date: 12/11/2025