

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-12112025-008

Gage ID PEA-45246 Gage S/N 45246 Description UT Thickness Gauge Operating Procedure: OP-121 Unit of Meas. inch Manufacturer Dakota Cal. Date 12/11/2025 Next Due 6/11/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 25-45% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15240 Job 14897

Standard ID	Uncertainty	Units	Type	V
.100 "	Minimum	Nominal	Maximum	0.102
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 5 Step Block	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No.	171846	
Std Due Date 8/21/2026				
Standard ID	Uncertainty	Units	Type	V
.200 "	Minimum	Nominal	Maximum	0.202
Limited Use? No	As Found	Accuracy	Fail Before	No
Ref Type 5 Step Block	As Left	Accuracy	Fail After	No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No.	171846	
Std Due Date 8/21/2026				
Standard ID	Uncertainty	Units	Type	V
.300 "	Minimum	Nominal	Maximum	0.302
Limited Use? No				

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214

Std Due Date 8/21/2026

Gage S/N 337214

NIST No. 171846

Calibrated By Chad Adams

Signature



Date: 12/11/2025

APPROVED*By Kayla Myers at 8:05 am, Dec 17, 2025*