Tubular Inspection Products
1551 Windway
Odessa, TX 79763
(432) 363-2200
PO Box 862
Odessa, TX 79760
Fax: (432) 363-2202

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	**Manufacturer**	Insize
Calibration Date	September 22, 2026	**TIP Procedure**	TIP-DGW01.1
Calibration Due Date	March 22, 2027	**Calibration Technician**	Freddy Seabourn
Serial Number	1504251346	**Order Number**	81796
Instrument	12" Digital Depth Gauge		

CALIBRATION RESULTS

Reference Size	Actual Measurement	Deviation	Tolerance	Result
1.000"	1.000"	+0.000"	±0.001"	**PASS**
2.000"	2.000"	+0.000"	±0.001"	**PASS**
3.000"	3.000"	+0.000"	±0.001"	**PASS**
4.000"	4.000"	+0.000"	±0.001"	**PASS**

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	**PASSED**

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Optical Comparator, Keyence IM-8030, S/N 4C0Q000412 (Cert. Due 01/28/2027)
- TIP Master Blocks 13232, 14002, 14340, 13207 (Cert. Due 01/28/2027)
- TIP Master Hygrometer, S/N 31874 / 1H2048MCR / 130070751 (Cert. Due 01/28/2027)

Calibration Technician

Freddy Seabourn

Issue Date: September 22, 2026

APPROVED
By Kayla Myers at 3:49 pm, Oct 30, 2026