



**Tubular Inspection
Products**

1551 Windway
Odessa, TX 79763
(432) 363-2200

PO Box 862
Odessa, TX 79760
Fax: (432) 363-2202

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	Manufacturer	R.B. Annis
Calibration Date	September 22, 2026	TIP Procedure	TIP-FIW01.1
Calibration Due Date	March 22, 2027	Calibration Technician	Freddy Seabourn
Serial Number	040626-01	Order Number	81796
Instrument	20-0-20 Field Indicator		

CALIBRATION RESULTS

Tolerance: $\pm 5\%$

Test	Tested Measurement	Actual Measurement
Positive Polarity Test	+20 G @ 12"	+20 G @ 12"
Zero Test	-20 G	-20 G
Negative Polarity Test	-20 G @ 12"	-20 G @ 12"

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	PASSED

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Coil, 6" AC Rectified Coil, S/N 100280 (Cert. Due 01/28/2027)
- TIP Master Gauss Meter, Insize 0501-MA01, S/N 5123041034 (Cert. Due 01/28/2027)
- TIP Master Hygrometer, S/N 31874 / 1H2048MCR / 130070751 (Cert. Due 01/28/2027)

Calibration Technician

Freddy Seabourn

Issue Date: September 22, 2026

APPROVED

By Kayla Myers at 1:39 pm, Sep 24, 2026