

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-12222025-002

Gage ID PEA-8415-3 Gage S/N 8415-3 Description 20-0-20 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 12/22/2025 Next Due 6/22/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11
Ref Type	Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No
Gage ID of Standard	Ref-09-1229						
Std Due Date	10/13/2026	Gage S/N	09-1229	NIST No.	114537		
Standard ID	+20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No
Gage ID of Standard	Ref-09-1229						
Std Due Date	10/13/2026	Gage S/N	09-1229	NIST No.	114537		
Standard ID	-10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11

Ref Type	Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	Gauss Meter	As Found	19	Accuracy	-1	Fail Before	No
		As Left	19	Accuracy	-1	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo

Signature



Date: 12/22/2025

APPROVED*By Kayla Myers at 10:55 am, Dec 22, 2025*