

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-06172025-014

<b>Gage ID</b> PEA-8415-3 <b>Gage S/N</b> 8415-3 <b>Description</b> 20-0-20 Gauss Meter <b>Operating Procedure:</b> OP-117 <b>Unit of Meas.</b> Gauss <b>Manufacturer</b> R. B. Annis Co., Inc. <b>Cal. Date</b> 6/17/2025 <b>Next Due</b> 12/17/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 +/- 2 deg F <b>Humidity</b> 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	+10	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9	<b>Nominal</b>	10	<b>Maximum</b>	11
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	10	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	10	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	Ref-09-1229						
<b>Std Due Date</b>	10/10/2025	<b>Gage S/N</b>	09-1229	<b>NIST No.</b>	114537		
<b>Standard ID</b>	+20	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	19	<b>Nominal</b>	20	<b>Maximum</b>	21
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	20	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	20	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	Ref-09-1229						
<b>Std Due Date</b>	10/10/2025	<b>Gage S/N</b>	09-1229	<b>NIST No.</b>	114537		
<b>Standard ID</b>	-10	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9	<b>Nominal</b>	10	<b>Maximum</b>	11

Ref Type	Gauss Meter	As Found	9	Accuracy	-1	Fail Before	No
		As Left	9	Accuracy	-1	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	Gauss Meter	As Found	19	Accuracy	-1	Fail Before	No
		As Left	19	Accuracy	-1	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo

Signature



Date: 6/17/2025

**APPROVED***By Kayla Myers at 3:24 pm, Jun 17, 2025*