

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04062026-011

<p>Gage ID PEA-040626-02 Gage S/N 040626-02 Description IR Thermometer Operating Procedure: OP-200 Unit of Meas. Degrees Fahrenheit Manufacturer Klein Tools Cal. Date 4/6/2026 Next Due 4/6/2027 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

As Found: Passed	As Left: Passed	
Heat Setting 1 Master: 218	As Found: 224	
Heat Setting 2 Master: 341	As Found: 349	
Heat Setting 3 Master: 452	As Found: 457	
Heat Setting 4 Master: 576	As Found: 583	
Heat Setting 5 Master: 684	As Found: 692	

APPROVED

By Kayla Myers at 10:03 am, Apr 14, 2026

Tolerance: +/- 7.5%

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
 Reference Standard: Fluke 62 Max IR Thermometer - Serial Number: 44070254WS
 NIST Number: Doc#171847 - Calibration Due Date: 6/17/26
 Measurements were achieved using a heating element at 5 different settings.

Calibrated By Chad Adams Signature  Date: 4/6/2026