

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06172025-013

Gage ID PEA-JZ23JG0051 Gage S/N JZ23JG0051 Description Pit Gauge Operating Procedure: OP-122 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 6/17/2025 Next Due 6/17/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	.10000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.09900	Nominal	0.10000	Maximum	0.10100
Ref Type	Mic Trac	As Found	0.10004	Accuracy	0.00004	Fail Before	No
		As Left	0.10004	Accuracy	0.00004	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	9/30/2025	Gage S/N	G141308	NIST No.	747V66-240502-00		
Standard ID	.20000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.19900	Nominal	0.20000	Maximum	0.20100
Ref Type	Mic Trac	As Found	0.20003	Accuracy	0.00003	Fail Before	No
		As Left	0.20003	Accuracy	0.00003	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	9/30/2025	Gage S/N	G141308	NIST No.	747V66-240502-00		
Standard ID	.30000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.29900	Nominal	0.30000	Maximum	0.30100

Ref Type	Mic Trac	As Found	0.30003	Accuracy	0.00003	Fail Before	No
		As Left	0.30003	Accuracy	0.00003	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 9/30/2025

Gage S/N G141308

NIST No. 747V66-240502-00

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40008	Accuracy	0.00008	Fail Before	No
		As Left	0.40008	Accuracy	0.00008	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 9/30/2025

Gage S/N G141308

NIST No. 747V66-240502-00

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50006	Accuracy	0.00006	Fail Before	No
		As Left	0.50006	Accuracy	0.00006	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 9/30/2025

Gage S/N G141308

NIST No. 747V66-240502-00

Calibrated By Taichi Daimo

Signature



Date: 6/17/2025

APPROVED*By Kayla Myers at 3:23 pm, Jun 17, 2025*