

3501 US Hwy 90 E  
Broussard, LA 70518

Date	S.O. No.
2/8/2024	12559

Name / Address
Peak NDT P.O. Box 671 Scott, LA 70583

Ship To
TSP 3501 US Hwy 90 E Broussard, LA 70518

P.O. No.	Rep	Job #
7		13109

Item	Description	Ordered	Rate	Amount
Test Weight Certif...	10 lb. Test Weight Calibration - S/N: 24498	1	30.00	30.00
			<b>Total</b>	\$30.00

Date:

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-02082024-000

<b>Gage ID</b> PEA-24498 <b>Gage S/N</b> 24498 <b>Description</b> 10 lb. Test Weight <b>Operating Procedure:</b> OP-118 <b>Unit of Meas.</b> lbs <b>Manufacturer</b> Parker <b>Cal. Date</b> 2/8/2024 <b>Next Due</b> 2/8/2027 <b>Cal. Freq.</b> 3.00 Years <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 12559 JOB 13109

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 7129341906

Test weight as found 10.11 lbs.

Test weight as left 10.11 lbs.

Calibration due date for scale is on 4/13/2024.

Scale is traceable to NIST by the following numbers:LA22-255, LA22-256, LA22-257

Calibrated By Taichi Daimo

Signature

Date: 2/8/2024