

TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04062026-003

<p>Gage ID PEA-180703022 Gage S/N 080703022 Description White Light Meter Operating Procedure: OP-130 Unit of Meas. Foot Candles Manufacturer General Cal. Date 4/6/2026 Next Due 10/6/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15815 Job 15231

	Master	As Found
High:	112.3	110.9
Low:	12.9	11.3

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND" AND "AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

Tolerance: +/-5% or +/-2 foot candles, whichever is greater.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Dual Light Meter - Serial Number: 1705427 Sensor: 1705430

NIST Number: Doc#171854 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams

Signature



Date: 4/6/2026

APPROVED
By Kayla Myers at 9:58 am, Apr 14, 2026