

SDM CALIBRATIONS LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185640	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	PEA-2802222223
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept:	WO# 070826-12	Mfg.:	Insize
		Cal. Date:	07/10/2026	Model:	Depth Gauge
		PO#:	VERBALCHRIS	Gage	12"

Depth Gauge Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.000000	+0.000000	+0.000000		+0.001500 / -0.001500
+0.500000	+0.499500	-0.000500		
+1.000000	+0.999500	-0.000500		
+6.000000	+5.999500	-0.000500		
+12.000000	+11.999500	-0.000500		

Comments

Procedure# SDM-49 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial#: 5.3155HD16

Gage Block Set 46pc., Starrett Model Grade B

Traceable to NIST via NAMAS certification no: Cert #

CN304-48449-711

Last / Next Cal Dates: 10/31/2023 -> 10/31/2026

Gage Status PASS

Next Calibration Due: 07/10/2027

Certified By: Mason Pellerin

Signature



This certificate is not valid unless all 1 page(s) are present

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 4:34 pm, Jul 13, 2026