

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-02182026-022

<p>Gage ID PEA-BR10LL0052 Gage S/N BR10LL0052 Description External Taper Gauge Operating Procedure: OP-132 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 2/18/2026 Next Due 8/18/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID .10000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.09900	Nominal 0.10000	Maximum 0.10100
Ref Type Mic Trac	As Found 0.09992	Accuracy -0.00008	Fail Before No
	As Left 0.09992	Accuracy -0.00008	Fail After No
Gage ID of Standard REF-G141308	Gage S/N G141308	NIST No. 747V66-240502-00	
Std Due Date 10/1/2026			
Standard ID .20000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.19900	Nominal 0.20000	Maximum 0.20100
Ref Type Mic Trac	As Found 0.19994	Accuracy -0.00006	Fail Before No
	As Left 0.19994	Accuracy -0.00006	Fail After No
Gage ID of Standard REF-G141308	Gage S/N G141308	NIST No. 747V66-240502-00	
Std Due Date 10/1/2026			
Standard ID .30000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.29900	Nominal 0.30000	Maximum 0.30100

Ref Type	Mic Trac	As Found	0.30012	Accuracy	0.00012	Fail Before	No
		As Left	0.30012	Accuracy	0.00012	Fail After	No

Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40015	Accuracy	0.00015	Fail Before	No
		As Left	0.40015	Accuracy	0.00015	Fail After	No

Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50009	Accuracy	0.00009	Fail Before	No
		As Left	0.50009	Accuracy	0.00009	Fail After	No

Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		

Calibrated By Taichi Daimo Signature  Date: 2/18/2026

APPROVED
By Kayla Myers at 1:39 pm, Feb 24, 2026