

1115 Garber Road
Broussard LA 70518

Company	Peak NDT Solutions	Doc #:	185626	Print#:	
Address	515A Verot School Road	OEM #:		Serial#:	PEA-13431-3
	Lafayette LA	Contact		Location	Calibration Lab
Zip:	70506	Dept	WO# 070826-12	Mfg.:	R.B. Annis
Page	1 of 1	Cal. Date	07/10/2026	Model	Magnetometer
		PO#:	VERBALCHRIS	Gage	20 Gauss

Gauss Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.000000	+0.000000	+0.000000		+0.400000 / -0.400000
+20.000000	+20.000000	+0.000000		
-20.000000	-20.000000	+0.000000		

Comments

Procedure# SDM-16 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 10-M-38
Helmholtz Coil (20 Gauss), R.B. Annis, Model 25B
Traceable to NIST via NAMAS certification no: Cert # 114537
Last / Next Cal Dates: 08/20/2024 -> 08/20/2026

Gage Status PASS

Next Calibration Due: 07/10/2027

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present
NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS
- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 4:35 pm, Jul 13, 2026