

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-07172025-016

Gage ID PEA-220942452 Gage S/N 220942452 Description Groove Mic Operating Procedure: OP-111 Unit of Meas. Inch Manufacturer Insize Cal. Date 7/17/2025 Next Due 1/17/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	.100 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.099	Nominal	0.100	Maximum	0.101
Ref Type	GAGE BLOCKS	As Found	0.100	Accuracy	0.000	Fail Before	No
		As Left	0.100	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
Standard ID	.300 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.299	Nominal	0.300	Maximum	0.301
Ref Type	GAGE BLOCKS	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.499	Nominal	0.500	Maximum	0.501

Ref Type	GAGE BLOCKS	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
Standard ID	.700 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.699	Nominal	0.700	Maximum	0.701
Ref Type	GAGE BLOCKS	As Found	0.700	Accuracy	0.000	Fail Before	No
		As Left	0.700	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
Standard ID	.900 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.899	Nominal	0.900	Maximum	0.901
Ref Type	GAGE BLOCKS	As Found	0.900	Accuracy	0.000	Fail Before	No
		As Left	0.900	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Calibrated By Taichi Daimo Signature  Date: 7/17/2025

APPROVED

By Kayla Myers at 9:58 am, Jul 18, 2025