

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-02182026-023

<p><b>Gage ID</b> PEA-BR10LL0051  <b>Gage S/N</b> BR10LL0051  <b>Description</b> Internal Taper Gage  <b>Operating Procedure:</b> OP-133  <b>Unit of Meas.</b> inch  <b>Manufacturer</b> Gagemaker  <b>Cal. Date</b> 2/18/2026  <b>Next Due</b> 8/18/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved Yes</b>  <b>Customer Info.</b> Peak NDT</p>
--	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	.10000"	<b>Uncertainty</b>		<b>Units</b>	<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.09900	<b>Nominal</b>	0.10000	<b>Maximum</b> 0.10100
<b>Ref Type</b>	Mic Trac	<b>As Found</b>	0.09993	<b>Accuracy</b>	-0.00007	<b>Fail Before</b> No
		<b>As Left</b>	0.09993	<b>Accuracy</b>	-0.00007	<b>Fail After</b> No
<b>Gage ID of Standard</b>	REF-G141308	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00	
<b>Std Due Date</b>	10/1/2026					

<b>Standard ID</b>	.30000"	<b>Uncertainty</b>		<b>Units</b>	<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.29900	<b>Nominal</b>	0.30000	<b>Maximum</b> 0.30100
<b>Ref Type</b>	Mic Trac	<b>As Found</b>	0.29991	<b>Accuracy</b>	-0.00009	<b>Fail Before</b> No
		<b>As Left</b>	0.29991	<b>Accuracy</b>	-0.00009	<b>Fail After</b> No
<b>Gage ID of Standard</b>	REF-G141308	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00	
<b>Std Due Date</b>	10/1/2026					

<b>Standard ID</b>	.50000"	<b>Uncertainty</b>		<b>Units</b>	<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.49900	<b>Nominal</b>	0.50000	<b>Maximum</b> 0.50100

Ref Type	Mic Trac	As Found	0.49991	Accuracy	-0.00009	Fail Before	No
		As Left	0.49991	Accuracy	-0.00009	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 10/1/2026

Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.70000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.69900	Nominal	0.70000	Maximum	0.70100
Ref Type	Mic Trac	As Found	0.69991	Accuracy	-0.00009	Fail Before	No
		As Left	0.69991	Accuracy	-0.00009	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 10/1/2026

Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.90000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.89900	Nominal	0.90000	Maximum	0.90100
Ref Type	Mic Trac	As Found	0.89988	Accuracy	-0.00012	Fail Before	No
		As Left	0.89988	Accuracy	-0.00012	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 10/1/2026

Gage S/N G141308 NIST No. 747V66-240502-00

Calibrated By Taichi Daimo Signature  Date: 2/18/2026

**APPROVED**  
 By Kayla Myers at 1:42 pm, Feb 24, 2026