TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01132025-010

Gage ID PEA-H06416 Gage S/N H06416 **Description** 5 Step Block

Operating OP-126

Procedure:

Unit of Meas. Inch

Manufacturer N/A

Cal. Date 1/13/2025 Next Due 1/13/2026

Cal. Freq. 1.00

Years

Location Lab

Approved Yes Customer Info. Peak NDT

20-55%

68 +/- 2 deg F

Environmental Conditions

Temperature

Humidity

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	0.100	•	Uncertainty			Units		Type	٧
Limited Use?	No		Minimum		0.098	Nominal	0.100	Maximum	0.102
Ref Type	6" Digital Caliper		As Found		0.100	Accuracy	0.000	Fail Before	No
			As Left		0.100	Accuracy	0.000	Fail After	No
Gage ID of Star	ndard	REF-08341639							
Std Due Date 10/10/2025		10/10/2025	Gage S/N	08341639 NIST No.		CN304-48449-711			
Standard ID	0.200'		Uncertainty			Units		Туре	V
Limited Use?	No		Minimum		0.198	Nominal	0.200	Maximum	0.202
Ref Type	6" Digital Caliper		As Found		0.200	Accuracy	0.000	Fail Before	No
			As Left		0.200	Accuracy	0.000	Fail After	No
Gage ID of Star	ndard	REF-08341639							
Std Due Date	10/10/2025		Gage S/N	08341639		NIST No.	CN304-48449-711		
Standard ID	0.300	ı	Uncertainty		·····	Units		Туре	V
Limited Use?	No		Minimum		0.298	Nominal	0.300	Maximum	0.302

Ref Type 6" Digital Caliper As Found 0.300 Accuracy 0.000 Fail Before No As Left 0.300 0.000 Fail After No Accuracy Gage ID of Standard REF-08341639 Std Due Date NIST No. 10/10/2025 Gage S/N 08341639 CN304-48449-711 Standard ID 0.400" Uncertainty Units Type V Limited Use? 0.398 0.400 No Minimum Nominal Maximum 0.402 **Ref Type** 6" Digital Caliper As Found 0.400 Accuracy 0.000 Fail Before No As Left 0.400 0.000 Fail After No Accuracy Gage ID of Standard REF-08341639 Std Due Date 10/10/2025 Gage S/N 08341639 NIST No. CN304-48449-711 Standard ID 0.500 " Units Type V Uncertainty Limited Use? No Minimum 0.498 Nominal 0.500 Maximum 0.502 Ref Type 6" Digital Caliper As Found 0.500 Accuracy 0.000 Fail Before No As Left 0.500 Accuracy 0.000 Fail After No Gage ID of Standard REF-08341639

Gage S/N 08341639

Signature

NIST No.

CN304-48449-711

Date:

1/13/2025

Page

2 of 2

APPROVED

Calibrated By Taichi Daimo

10/10/2025

Gage ID: PEA-H06416

Std Due Date

By Kayla Myers at 9:39 am, Jan 14, 2025