

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	185642	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	21783
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 070826-12	Mfg.:	Dakota
		Cal Date:	07/10/2026	Model	Ultrasonic Thickness Gauge
		PO#:		Gage	.025-19.999"

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.250000	+0.251000	+0.001000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.500000	+0.501000	+0.001000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+0.750000	+0.751000	+0.001000		+0.001000 / -0.001000

Thickness Readings As Found

Master	Actual	Deviation	Serial	Tol
+1.000000	+1.001000	+0.001000		+0.001000 / -0.001000

Comments

Procedure# SDM-15 Rev.1
Accuracy +/- .001"

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 422711

4-Step Test Block, TSP, Model 1018 Steel

Traceable to NIST via NAMAS certification no: Cert# 80602

Last / Next Cal Dates 12/21/2023 → 12/21/2026

Gage Status PASS

Next Calibration Due: 01/10/2027

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibration certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-65%)

APPROVED

By Kayla Myers at 4:39 pm, Jul 13, 2026