

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-02182026-020

| | |
|---|--|
| <p>Gage ID PEA-22254 Gage S/N 22254 Description UT Thickness Gage Operating Procedure: OP-121 Unit of Meas. inch Manufacturer Dakota Ultrasonics Cal. Date 2/18/2026 Next Due 8/18/2026 Cal. Freq. 6.00 Months Location Lab</p> | <p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p> |
|---|--|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

| | | | | | | | |
|----------------------------|--------------|--------------------|--------|-----------------|--------|--------------------|-------|
| Standard ID | 0.250" | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 0.248 | Nominal | 0.250 | Maximum | 0.252 |
| Ref Type | 8 Step Block | As Found | 0.248 | Accuracy | -0.002 | Fail Before | No |
| | | As Left | 0.248 | Accuracy | -0.002 | Fail After | No |
| Gage ID of Standard | Ref-419719 | | | | | | |
| Std Due Date | 12/11/2026 | Gage S/N | 419719 | NIST No. | 176480 | | |
| Standard ID | 0.750" | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 0.748 | Nominal | 0.750 | Maximum | 0.752 |
| Ref Type | 8 Step Block | As Found | 0.750 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 0.750 | Accuracy | 0.000 | Fail After | No |
| Gage ID of Standard | Ref-419719 | | | | | | |
| Std Due Date | 12/11/2026 | Gage S/N | 419719 | NIST No. | 176480 | | |
| Standard ID | 1.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 0.998 | Nominal | 1.000 | Maximum | 1.002 |

| | | | | | | | |
|----------|--------------|----------|-------|----------|-------|-------------|----|
| Ref Type | 8 Step Block | As Found | 1.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 1.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard Ref-419719
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

| | | | | | | | |
|--------------|--------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 1.500 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 1.498 | Nominal | 1.500 | Maximum | 1.502 |
| Ref Type | 8 Step Block | As Found | 1.500 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 1.500 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard Ref-419719
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

| | | | | | | | |
|--------------|--------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 2.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 1.998 | Nominal | 2.000 | Maximum | 2.002 |
| Ref Type | 8 Step Block | As Found | 2.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 2.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard Ref-419719
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

Calibrated By Taichi Daimo Signature  Date: 2/18/2026