

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01072026-016

<p>Gage ID PEA-2116001923 Gage S/N 2116001923 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 1/7/2026 Next Due 7/7/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 2234240 was also used during this calibration.
 Calibration due date for black light is on 06/17/2026

Black light is traceable to NIST by the following numbers: 263546, 263547

Standard ID 12"	Uncertainty	Units	Type V
Limited Use? No	Minimum 6587	Nominal 6933	Maximum 7279
Ref Type Long Wave	As Found 7115	Accuracy 182	Fail Before No
	As Left 7115	Accuracy 182	Fail After No
Gage ID of Standard REF-1944278	Gage S/N 1944278/1944 280	NIST No. 263546, 263547	
Std Due Date 6/17/2026			

Standard ID 15"	Uncertainty	Units	Type V
Limited Use? No	Minimum 4521	Nominal 4758	Maximum 4995
Ref Type Long Wave	As Found 4936	Accuracy 178	Fail Before No
	As Left 4936	Accuracy 178	Fail After No
Gage ID of Standard REF-1944278	Gage S/N 1944278/1944 280	NIST No. 263546, 263547	
Std Due Date 6/17/2026			

Calibrated By Taichi Daimo

Signature 

Date: 1/7/2026

APPROVED

By Kayla Myers at 3:50 pm, Jan 07, 2026