

Company	Peak NDT Solutions	Doc #:	185644	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	240302879
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 070826-12	Mfg.:	General
		Cal Date	07/10/2026	Model	Visible Light Meter
		PO#:		Gage	0-2,000 fc

Visible Light Readings As Found @ 15" (High)

Master	Actual	Deviation	Serial	Tol
+350.000000	+340.000000	-10.000000		+10.500000 / -10.500000

Visible Light Readings As Found @ 15" (Low)

Master	Actual	Deviation	Serial	Tol
+30.000000	+30.900000	+0.900000		+0.900000 / -0.900000

Comments

Procedure# SDM-18 Rev.1
Accuracy +/- 3% or reading 5% @ high range

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 2038976 & 2038979

Radiometer/Photometer, Spectroline Model XRP-3000

Traceable to NIST via NAMA certification no: Cert# 272637,
272638, 272639, 272640

Last / Next Cal Dates 10/02/2025 -> 10/02/2026

Gage Status PASS**Next Calibration Due: 01/10/2027**

Certified By: Mason Pellerin

Signature



This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees, Humidity 20-55%)

APPROVED

By Kayla Myers at 4:39 pm, Jul 13, 2026