

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-07172025-008

<b>Gage ID</b> PEA-210700984 <b>Gage S/N</b> 210700984 <b>Description</b> Visible Light Meter <b>Operating Procedure:</b> OP-130 <b>Unit of Meas.</b> Foot Candles <b>Manufacturer</b> General <b>Cal. Date</b> 7/17/2025 <b>Next Due</b> 1/17/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	97.7	95.7
Low	35.5	33.8

**APPROVED****By Kayla Myers at 9:10 am, Jul 18, 2025**

Readings As Left (foot candles)

	Master	Actual
High	97.7	95.7
Low	35.5	33.8

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater  
Cal due date for light meter is 06/17/2026  
NIST Traceable through : 263546, 263547

Calibrated By Taichi Daimo

Signature

Date: 7/17/2025